


Sterilizer and QuickClean Shutdown Procedures

If you need to shut down your sterilizers and/or QuickClean™ units for an extended period of time, please follow the simple steps on the next page.



M9 and M11 Steam Sterilizers

Caution: Before cleaning, make sure your sterilizer unit is cooled to room temperature.

STEP 1: Wash the exterior of sterilizer using a cloth or wipe containing a quaternary amine disinfectant. Do not use metal brushes, steel wool or other abrasives to clean the unit. Do not use solutions containing alcohol or bleach.

STEP 2: Rinse the sterilizer exterior with distilled water* and let dry completely.

STEP 3: Open the sterilizer door. Clean the gaskets and mating surfaces with a damp cloth.

STEP 4: Using the drain tube mounted on the front of the unit, drain the water from the reservoir. With assistance, tilt the unit to ensure the reservoir has been drained as completely as possible. Note: This step is unnecessary if the sterilizer is connected to a direct-to-drain thermal reduction system.

STEP 5: Remove all trays, racks and plates from the interior of the sterilizer. Wash these items and the inside of the sterilizer chamber with mild soap and water. Rinse with distilled water* after washing.

STEP 6: Let chamber air dry completely before closing the door.

M3 Sterilizers

Caution: Before cleaning, make sure your sterilizer unit is cooled to room temperature.

STEP 1: Drain water from the reservoir.

STEP 2: Prime the pump for 5-10 seconds using the following process:

- A.** Put the unit in Diagnostic Mode as described in the User's Guide.
- B.** Press the **Start** button to start the pump priming. The unit will automatically perform a pre-programmed cycle, and when the cycle is finished, the 2nd line of the display will show "PRIMING COMPLETE."
- C.** Press the **Start** button to return to the Diagnostic Menu.
- D.** Turn the power switch "Off" to exit Diagnostic mode.

STEP 3: With the power switch off, press and hold the <P> button while turning on the power switch.

STEP 4: Release the <P> button when the display reads "SERVICE DIAGNOSTICS."

STEP 5: Press and hold the **Start** button until "KEY TEST" appears on the display.

STEP 6: Press the **Stop** button 4 times. When the display shows "BOILER TEST," press the **Start** button to power on the boiler.

STEP 7: Once the boiler starts, wait 30 seconds and then turn the power switch off.

STEP 8: With assistance, cautiously tip the unit forward to fully empty the reservoir.

STEP 9: Remove the door/tray and dry the tray and inside of the chamber.

STEP 10: Reinsert the door/tray in the chamber.

STEP 11: Drain and wipe out the external condensing tank.

QuickClean™ Units

STEP 1: Drain the cleaning solution from the tank.

STEP 2: Remove the baskets and wipe down the inside of tank with soap and distilled water*.

STEP 3: Clean and rinse the baskets with soap and distilled water* and let air dry.

STEP 4: Replace the baskets when dry.

*Water, when referenced for washing or rinsing, is defined as distilled water or water that meets the water purity specifications found in the [M9/M11](#), [M3](#) and [QuickClean](#) User Guides.