

QuickVue® RSV Quiz

Circle the correct answer

This quiz is an educational tool intended to assist facilities in evaluating their operators' understanding of the QuickVue RSV test procedure. This quiz is not intended to be used as sole evidence of operator training or competency. Facilities are responsible for ensuring the quality of the testing performed by their operators. When testing controls or patient specimens, follow the current package insert instructions and/or procedure card included in the kit.

1. How often does the manufacturer recommend that external controls be run on the kit?

- a) Once every 20 tests
- b) Once per each new shipment of kits
- c) Every first sample tested each day
- d) Once per each untrained operator, once for each new shipment of kits-provided that each different lot in a shipment is tested

2. At what temperature do the kits need to be stored?

- a) 2-8°C (refrigerated)
- b) 15-30°C (room temp.)
- c) 2-30°C (refrigerated or room temp.)
- d) -20°C (frozen)

3. For proper test performance, which swab should be used when collecting a nasopharyngeal swab specimen?

- a) The swabs provided in the kit
- b) Any type of swabs
- c) Nasal swabs
- d) Cotton tipped swabs

4. What is the read time for this kit?

- a) 10 minutes
- b) 15 minutes
- c) 12 minutes
- d) 20 minutes

5. Can the test be read after the read time?

- a) No
- b) Yes, up to 20 minutes after the read time
- c) Yes, up to 5 minutes after the read time
- d) Yes, any time after the read time

6. How much Extraction Reagent should I add into the tube?

- a) All of the reagent in the bottle
- b) Up to the fill line
- c) None
- d) 8 drops

7. What is the incubation time when the sample is in the Extraction Reagent?

- a) There is no incubation time
- b) 15 minutes
- c) 1 minute
- d) 1-2 minutes

8. What are the age limitations of the QuickVue RSV test?

- a) There are none
- b) 5 years and under
- c) 18 years and under
- d) 21 years and under

9. What specimen types can be used when running this test?

- a) Nasopharyngeal swab and nasopharyngeal aspirate
- b) Nasopharyngeal swab, nasopharyngeal aspirate, nasal wash, and nasopharyngeal wash
- c) Nasopharyngeal swab, nasopharyngeal aspirate, and nasal wash
- d) Nasopharyngeal swab and nasal swab

10. Which viral transport media are not compatible with this kit?

- a) M4-RT
- b) Multitrans Media and UTM
- c) M4-3 and Amies
- d) M4-3 and M4-RT

QuickVue® RSV Quiz

Circle the correct answer

This quiz is an educational tool intended to assist facilities in evaluating their operators' understanding of the QuickVue RSV test procedure. This quiz is not intended to be used as sole evidence of operator training or competency. Facilities are responsible for ensuring the quality of the testing performed by their operators. When testing controls or patient specimens, follow the current package insert instructions and/or procedure card included in the kit.

1. How often does the manufacturer recommend that external controls be run on the kit?

- a) Once every 20 tests
- b) Once per each new shipment of kits
- c) Every first sample tested each day
- d) Once per each untrained operator, once for each new shipment of kits-provided that each different lot in a shipment is tested

2. At what temperature do the kits need to be stored?

- a) 2-8°C (refrigerated)
- b) 15-30°C (room temp.)
- c) 2-30°C (refrigerated or room temp.)
- d) -20°C (frozen)

3. For proper test performance, which swab should be used when collecting a nasopharyngeal swab specimen?

- a) The swabs provided in the kit
- b) Any type of swabs
- c) Nasal swabs
- d) Cotton tipped swabs

4. What is the read time for this kit?

- a) 10 minutes
- b) 15 minutes
- c) 12 minutes
- d) 20 minutes

5. Can the test be read after the read time?

- a) No
- b) Yes, up to 20 minutes after the read time
- c) Yes, up to 5 minutes after the read time
- d) Yes, any time after the read time

6. How much Extraction Reagent should I add into the tube?

- a) All of the reagent in the bottle
- b) Up to the fill line
- c) None
- d) 8 drops

7. What is the incubation time when the sample is in the Extraction Reagent?

- a) There is no incubation time
- b) 15 minutes
- c) 1 minute
- d) 1-2 minutes

8. What are the age limitations of the QuickVue RSV test?

- a) There are none
- b) 5 years and under
- c) 18 years and under
- d) 21 years and under

9. What specimen types can be used when running this test?

- a) Nasopharyngeal swab and nasopharyngeal aspirate
- b) Nasopharyngeal swab, nasopharyngeal aspirate, nasal wash, and nasopharyngeal wash
- c) Nasopharyngeal swab, nasopharyngeal aspirate, and nasal wash
- d) Nasopharyngeal swab and nasal swab

10. Which viral transport media are not compatible with this kit?

- a) M4-RT
- b) Multitrans Media and UTM
- c) M4-3 and Amies
- d) M4-3 and M4-RT